

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420

Email: info@geneabiomedx.com

Website: geneabiomedx.com

## Product Details

Product Description	GAVI TIP and SEAL
Product Code	GAVI-TIP-20
Lot Number	TS0049-001
Manufacture Date	2024-02-12
Expiry date	2024-11-20

## Product Specification

Requirement	Specification	Result
MEA	≥ 80% of 2PN's developed to Blastocyst stage by 96 hrs	Pass
Endotoxin	< 20EU/Device	Pass
Sterility	No Growth	Pass

This product complies with approved and released procedures within the Genea Biomedx EN/ISO 13485 compliant Quality Management System. This product and all its identified Part Numbers and Revisions conform to approved product specifications retained within the QMS. Where cited, lot numbers indicate traceable items within the QMS.

## Product Passed and Released by Quality Assurance

Signature:

Electronically signed by: Elliot Atchison  
Reason: Approval  
Date: May 13, 2024 13:34 GMT+10

Name: Elliot Atchison

Job Title: Manufacturing QA Team Leader

Date: 13 May. 2024