Certificate of Analysis



Phone: Email: Website:

+61 2 8228 2719 info@geneabiomedx.com geneabiomedx.com

Product Details

| Product description: | GAVI Medium Cartridge |
|----------------------|-----------------------|
| Product code | GAVI-MED-20 |
| Lot number: | J000006341 |
| Expiry date: | 21 Nov. 2024 |

Intended Use

Gavi Medium Cartridge is intended for use within the Gavi instrument (GAVI-INS-01) for the vitrification of human embryos for Assisted Reproductive Technology (ART) procedures.

GAVI-MED-20 comprises of Gavi Solution 1 containing 8 % DMSO and Gavi Solution 2 containing 16 % DMSO.

Product Analysis

| A | Specification | | Desult |
|--------------------|--|---------------------|--------|
| Assay | Solution 1 | Solution 2 | Result |
| pH at 25 °C in air | 7.3 – 7.5 | | Pass |
| MEA | MEA (1-cell): ≥ 80 % developed to blastocyst stage or beyond by 96 hours | | Pass |
| Endotoxin | <0.40 EU/mL | | Pass |
| Sterility | No Growth | | Pass |
| HSA Analysis | 12.60 – 21.00 mg/mL | 10.13 – 16.88 mg/mL | Pass |

Product passed and released by Quality

| Signature: | FMODERON E | Electronically signed by: Eila McGregor Reason: QA Approval Date: May 23, 2024 12:45 GMT+10 |
|------------|-----------------------------|---|
| Name: | Eila McGregor | |
| Job Title: | Quality Assurance Associate | |
| Date: | 23 May. 2024 | |