

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product description:	Blastocyst Medium
Product code	BLM-20
Lot number:	J000006387
Expiry date:	22 Jan. 2025

Intended Use

Blastocyst Medium is used to provide necessary nutrients for embryo development *in vitro* from the cleavage stage to the blastocyst stage. The embryos may remain in this solution for up to 3 days prior to being transferred to the uterus.

USA only: Gems Blastocyst Medium is for *in vitro* culture of embryos from the cleavage stage to the blastocyst stage of development. This medium can be used for transfer of blastocyst stage embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO ₂	7.25 – 7.45	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.4 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.8 – 6.2 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Revathy Veera Raghavalu
Reason: QA Approved
Date: Feb 29, 2024 16:03 GMT+11

Name: Revathy Veera Raghavalu

Job Title: QA & RA Associate

Date: 29 Feb. 2024