

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420  
Email: info@geneabiomedx.com  
Website: geneabiomedx.com

## Product Details

**Product description:** Sperm Medium  
**Product code:** SPM-50  
**Lot number:** J000006453  
**Expiry date:** 18 Feb. 2025

## Intended Use

Sperm Medium is used to wash and resuspend sperm for the insemination step in IUI, IVF or in diagnostic washing, optimised for storage in a 6% CO<sub>2</sub> incubator.

## Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO <sub>2</sub>	7.30 – 7.50	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.4 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.5 – 12.5 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

**Signature:**

**Name:** Elliot Atchison

**Job Title:** Manufacturing QA Team Leader

**Date:** 25 Mar. 2024