

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420

Email: [info@geneabiomedx.com](mailto:info@geneabiomedx.com)

Website: [geneabiomedx.com](http://geneabiomedx.com)

## Product Details

Product description:	Sperm Medium
Product code	SPM-50
Lot number:	J000006487
Expiry date:	05 Mar. 2025

## Intended Use

Sperm Medium is used to wash and resuspend sperm for the insemination step in IUI, IVF or in diagnostic washing, optimised for storage in a 6% CO<sub>2</sub> incubator.

## Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO <sub>2</sub>	7.30 – 7.50	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Sperm Toxicity Test	Sperm Survival Tested	Pass
Endotoxin	<0.4 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	7.5 – 12.5 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

Signature:

Electronically signed by: Lalitha Kalyanaraman  
Reason: QA Approval  
Date: Apr 12, 2024 12:05 GMT+10

Name:

Lalitha Kalyanaraman

Job Title:

QA Associate

Date:

12 Apr. 2024