

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420

Email: [info@geneabiomedx.com](mailto:info@geneabiomedx.com)

Website: [geneabiomedx.com](http://geneabiomedx.com)

## Product Details

**Product description:** Hyaluronidase  
**Product code:** HYA-01  
**Lot number:** J000006515  
**Expiry date:** 31 May. 2024

## Intended Use

Used for human oocyte cumulus cell removal.

## Product Analysis

Assay	Specification	Result
MEA	MEA (1-cell): $\geq 80\%$ developed to blastocyst stage or beyond by 96 hours	Pass

## Product passed and released by Quality

**Signature:**   
*Electronically signed  
by: Eila McGregor  
Reason: QA Approval  
Date: Apr 2, 2024  
09:58 GMT+11*

**Name:** Eila McGregor

**Job Title:** Quality Assurance Associate

**Date:** 02 Apr. 2024