# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420
Email: info@geneabiomedx.com
Website: geneabiomedx.com

### **Product Details**

**Product description:** Sperm Wash Gradient Set

Product code SWG-01
Lot number: J000006519
Expiry date: 17 Mar. 2025

## **Intended Use**

Sperm Wash Gradient Set is used to separate sperm from seminal plasma as well as separating highly motile sperm in preparation for insemination.

# **Product Analysis**

Assay	Specification		Result
	45% gradient	90% gradient	Result
pH at 25 °C in air	7.30 – 7.50		Pass
Osmolality	285 - 295 mOsm/kg	280 - 290 mOsm/kg	Pass
Sperm Toxicity Test	Sperm Survival Tested		Pass
Endotoxin	<0.4 EU/mL		Pass
Sterility	No Growth		Pass
HSA Analysis	7.5 – 12.5 mg/mL		Pass
Gentamicin	0.0075 – 0.0125 mg/mL		Pass

## Product passed and released by Quality

Signature: K. Lalitu

Electronically signed by: Lalitha Kalyanaraman Reason: QA Approval Date: Apr 16, 2024 13:17 GMT+10

Name: Lalitha Kalyanaraman

Job Title: QA Associate

**Date:** 16 Apr. 2024

QFRM221-09 Page 1 of 1