

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Hyaluronidase
Product code: HYA-01
Lot number: J000006648
Expiry date: 09 Aug. 2024


Intended Use

Used for human oocyte cumulus cell removal.

Product Analysis

Assay	Specification	Result
MEA	MEA (1-cell): $\geq 80\%$ developed to blastocyst stage or beyond by 96 hours	Pass

Product passed and released by Quality

Signature: 
Name: Eila McGregor
Job Title: Quality Assurance Associate
Date: 06 Jun. 2024