

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Oocyte Retrieval Buffer
Product code: ORB-50
Lot number: J000006820
Expiry date: 14 Aug. 2025

Intended Use

Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

| Assay | Specification | Result |
|--------------------|--|--------|
| pH at 25 °C in air | 7.3 – 7.5 | Pass |
| Osmolality | 280 – 290 mOsm/kg | Pass |
| MEA | MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours | Pass |
| Endotoxin | <0.40 EU/mL | Pass |
| Sterility | No Growth | Pass |
| Gentamicin | 0.0075 – 0.0125 mg/mL | Pass |

Product passed and released by Quality

Signature: 
Name: Eila McGregor
Job Title: Quality Assurance Associate
Date: 18 Sep. 2024