

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719  
Email: info@geneabiomedx.com  
Website: geneabiomedx.com

## Product Details

**Product description:** Oocyte Retrieval Buffer  
**Product code:** ORB-50  
**Lot number:** J000006890  
**Expiry date:** 03 Sep. 2025

## Intended Use


Gems Oocyte Retrieval Buffer is used in the removal of oocytes from ovarian follicles.

**Caution:** Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

## Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	280 – 290 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

**Signature:**   
**Name:** Eila McGregor  
**Job Title:** Quality Assurance Associate  
**Date:** 14 Oct. 2024