

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420  
Email: info@geneabiomedx.com  
Website: geneabiomedx.com

## Product Details

**Product description:** Hyaluronidase  
**Product code:** HYA-01  
**Lot number:** J000006245  
**Expiry date:** 21 Dec. 2023

## Intended Use

Used for human oocyte cumulus cell removal.

## Product Analysis

Assay	Specification	Result
MEA	MEA (1-cell): $\geq 80\%$ developed to blastocyst stage or beyond by 96 hours	Pass

## Product passed and released by Quality

**Signature:**

Electronically signed by: Elliot Atchison  
Reason: approval  
Date: Oct 17, 2023 10:32 GMT+11

**Name:** Elliot Atchison  
**Job Title:** Manufacturing QA Team Leader  
**Date:** 17 Oct. 2023