

# Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420

Email: [info@geneabiomedx.com](mailto:info@geneabiomedx.com)

Website: [geneabiomedx.com](http://geneabiomedx.com)

## Product Details

Product description:	Blastocyst Medium
Product code	BLM-50
Lot number:	J000006240
Expiry date:	08 Oct. 2024

## Intended Use

Blastocyst Medium is used to provide necessary nutrients for embryo development *in vitro* from the cleavage stage to the blastocyst stage. The embryos may remain in this solution for up to 3 days prior to being transferred to the uterus.

**USA only:** Gems Blastocyst Medium is for *in vitro* culture of embryos from the cleavage stage to the blastocyst stage of development. This medium can be used for transfer of blastocyst stage embryos into the uterus.

**Caution:** Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

## Product Analysis

Assay	Specification	Result
pH at 37°C with 6 % CO <sub>2</sub>	7.25 – 7.45	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.4 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.8 – 6.2 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

## Product passed and released by Quality

Signature:  Electronically signed by:  
Eila McGregor  
Reason: QA Approval  
Date: Jan 3, 2024 10:11  
GMT+11

Name: Eila McGregor

Job Title: Quality Assurance Associate

Date: 03 Jan. 2024