

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 9229 6420

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product description:	Cleavage Medium
Product code	CLM-50
Lot number:	J000006460
Expiry date:	21 Feb. 2025

Intended Use

Cleavage Medium is used to provide necessary nutrients for embryo development *in vitro*. The embryos can remain in this solution for up to 2 days prior to being transferred to the uterus or grown for up to another 3 days in Blastocyst Medium (BLM-20, BLM-50).

USA only: Gems Cleavage Medium is for *in vitro* culture of embryos following fertilisation to the 4-8 cell stage. This medium can be used for transfer of cleavage stage embryos into the uterus.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 37 °C with 6% CO ₂	7.30 – 7.50	Pass
Osmolality	285 – 295 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.4 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	3.8 – 6.2 mg/mL	Pass
Gentamicin	0.0075 – 0.0125 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Lalitha Kalyanaraman
Reason: QA Approval
Date: Mar 25, 2024 16:06 GMT+11

Name:

Lalitha Kalyanaraman

Job Title:

QA Associate

Date:

25 Mar. 2024