

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719

Email: info@geneabiomedx.com

Website: geneabiomedx.com

Product Details

Product description: Hyaluronidase
Product code: HYA-01
Lot number: J000006601
Expiry date: 05 Jul. 2024

Intended Use

Used for human oocyte cumulus cell removal.

Product Analysis

Assay	Specification	Result
MEA	MEA (1-cell): $\geq 80\%$ developed to blastocyst stage or beyond by 96 hours	Pass

Product passed and released by Quality

Signature: 
Electronically signed
by: Eila McGregor
Reason: QA Approval
Date: May 1, 2024
13:35 GMT+10

Name: Eila McGregor

Job Title: Quality Assurance Associate

Date: 01 May. 2024