

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: Vitrification Set
Product code: VIT-01
Lot number: J000006652
Expiry date: 22 May. 2025

Intended Use

Gems Vitrification Set is used for the vitrification of human embryos for Assisted Reproductive Technology (ART) procedures.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Set Contents

The Gems Vitrification set comprises of Vitsol 1, Vitsol 2 and Vitsol 3.

Product Analysis

Assay	Specification			Result
	Vitsol 1	Vitsol 2	Vitsol 3	
pH at 25 °C in air	7.3 – 7.5			Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours			Pass
Endotoxin	<0.40 EU/mL		<0.05 EU/mL	Pass
Sterility	No Growth			Pass
HSA Analysis	13.70 – 22.90 mg/mL	12.20 – 20.30 mg/mL		Pass

Product passed and released by Quality

Signature:

Electronically signed by: Elliot Atchison
Reason: Approval
Date: Jul 10, 2024 11:37 GMT+10

Name: Elliot Atchison

Job Title: Manufacturing QA Team Leader

Date: 10 Jul. 2024