

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: VitBase
Product code: VBS-20
Lot number: J000006704
Expiry date: 24 Jun. 2025

Intended Use

VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment.

USA only: Gems VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment during embryo vitrification and warming procedures. The embryos can be placed in this medium for a maximum of 10 minutes.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

| Assay | Specification | Result |
|--------------------|--|--------|
| pH at 25 °C in air | 7.3 – 7.5 | Pass |
| Osmolality | 295 – 305 mOsm/kg | Pass |
| MEA | MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours | Pass |
| Endotoxin | <0.40 EU/mL | Pass |
| Sterility | No Growth | Pass |
| HSA Analysis | 15.00 – 25.00 mg/mL | Pass |

Product passed and released by Quality

Signature:  Electronically signed by:
Eila McGregor
Reason: QA Approval
Date: Jul 24, 2024
15:16 GMT+10

Name: Eila McGregor
Job Title: Quality Assurance Associate
Date: 24 Jul. 2024