

Certificate of Analysis



Level 2, 321 Kent Street, Sydney NSW 2000, Australia

Phone: +61 2 8228 2719
Email: info@geneabiomedx.com
Website: geneabiomedx.com

Product Details

Product description: VitBase
Product code: VBS-20
Lot number: J000006893
Expiry date: 27 Aug. 2025

Intended Use

VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment.

USA only: Gems VitBase is a HEPES buffered medium that is used to maintain embryos for a short period of time in a non-gassed environment during embryo vitrification and warming procedures. The embryos can be placed in this medium for a maximum of 10 minutes.

Caution: Federal law (USA) restricts this device to sale by or on the order of a physician or practitioner trained in its use.

Product Analysis

Assay	Specification	Result
pH at 25 °C in air	7.3 – 7.5	Pass
Osmolality	295 – 305 mOsm/kg	Pass
MEA	MEA (1-cell): ≥80% developed to blastocyst stage or beyond by 96 hours	Pass
Endotoxin	<0.40 EU/mL	Pass
Sterility	No Growth	Pass
HSA Analysis	15.00 – 25.00 mg/mL	Pass

Product passed and released by Quality

Signature:

Electronically signed by: Lalitha Kalyanaraman
Reason: QA Approval
Date: Sep 26, 2024 10:54 GMT+10

Name: Lalitha Kalyanaraman

Job Title: QA Associate

Date: 26 Sep. 2024